City Research Online

City, University of London Institutional Repository

Citation: Cutts, S., Burls, A., Akinlabi, K., Hurst, J. R., Mandalia, S. & Mansell, S. K. (2018). A randomised crossover trial investigating the effect of a portable positive pressure ventilation device on exercise tolerance in patients with COPD. *European Respiratory Journal*, 52(Supp62), PA2054. doi: 10.1183/13993003.congress-2018.PA2054

This is the accepted version of the paper.

This version of the publication may differ from the final published version.

Permanent repository link: <https://openaccess.city.ac.uk/id/eprint/21037/>

Link to published version: <https://doi.org/10.1183/13993003.congress-2018.PA2054>

Copyright: City Research Online aims to make research outputs of City, University of London available to a wider audience. Copyright and Moral Rights remain with the author(s) and/or copyright holders. URLs from City Research Online may be freely distributed and linked to.

Reuse: Copies of full items can be used for personal research or study, educational, or not-for-profit purposes without prior permission or charge. Provided that the authors, title and full bibliographic details are credited, a hyperlink and/or URL is given for the original metadata page and the content is not changed in any way.

A randomised crossover trial investigating the effect of a portable positive pressure ventilation device on exercise tolerance in patients with COPD

Background: Dyspnoea is a common symptom in COPD and can lead to progressive decline in exercise capacity. Non-Invasive Ventilation has been shown to improve symptoms and exercise capacity but is clinically impractical due to logistical constraints. A portable ventilation device (Vitabreath) may overcome these barriers and be a useful adjunct therapy.

Aim: To investigate the effect of Vitabreath on exercise capacity in patients with COPD.

Methods: Randomised crossover design; 12 participants with COPD ($FEV_1\%$ 45 ± 15) performed 3 Six-Minute Walk Tests using i) Vitabreath, ii) Threshold Positive Expiratory Pressure (PEP), or iii) no device. Primary outcome: six-minute walk distance (6MWD). Secondary outcomes: changes in heart rate, oxygen saturations (SpO_2), dyspnoea, and lower limb (LL) fatigue, and recovery time of each variable.

Results: Mean 6MWD was less using Vitabreath compared to no device ($p=0.01$). Use of Vitabreath resulted in a smaller change in dyspnoea ($p=0.008$) and LL fatigue scores ($p=0.02$), and a faster LL recovery time ($p=0.01$) compared to Threshold PEP. SpO_2 recovery time was faster using Vitabreath compared to both Threshold PEP ($p=0.008$) and no device ($p=0.03$). Parametric data presented as mean $\pm$ SD, Non-parametric data presented as median (IQR).

Conclusion: The data suggest no benefit in using the Vitabreath in improving exercise capacity.

	Vitabreath	Threshold PEP	No device	ANOVA p=
6MWD (m)	417 $\pm$ 50	430 $\pm$ 67	465 $\pm$ 71	0.01
Change in Variable				
Dyspnoea (Borg)	2 (1)	3 (1)	3 (2)	0.03
LL fatigue (Borg)	0 (2)	2 (1)	1 (1)	0.03
Recovery time (seconds)				
LL fatigue	0 (120)	120 (180)	120 (180)	0.014
SpO_2	28 (62)	68 (113)	83 (83)	0.003