



City Research Online

City St George's, University of London

Citation: Kyriacou, P. A., Charlton, P. H., Al-Halawani, R. & Shelley, K. H. (2023). Inaccuracy of pulse oximetry with dark skin pigmentation: clinical implications and need for improvement. *British Journal of Anaesthesia*, 130(1), e33-e36. doi: 10.1016/j.bja.2022.03.011

This is the accepted version of the paper.

This version of the publication may differ from the final published version. To cite this item please consult the publisher's version.

Permanent repository link: <https://openaccess.city.ac.uk/id/eprint/28501/>

Link to published version: <https://doi.org/10.1016/j.bja.2022.03.011>

Copyright and Reuse: Copyright and Moral Rights remain with the author(s) and/or copyright holders. Copies of full items can be used for personal research or study, educational, or not-for-profit purposes without prior permission or charge, unless otherwise indicated, provided that the authors, title and full bibliographic details are credited, a hyperlink and/or URL is given for the original metadata page and the content is not changed in any way. For full details of reuse please refer to [City Research Online policy](#).

The effect of skin pigmentation on pulse oximeter accuracy

Panicos A. Kyriacou¹, Peter H. Charlton^{1,2}, Raghda Al-Halawani¹ and Kirk H. Shelley³

¹Research Centre for Biomedical Engineering, City, University of London, London, UK

²Department of Public Health and Primary Care, University of Cambridge, Cambridge, UK

³Department of Anaesthesiology, Yale University, New Haven, CT, USA.

Corresponding author: Panicos A. Kyriacou (P.Kyriacou@city.ac.uk)

Summary

This editorial is driven by recent reports highlighting potential inaccuracies of pulse oximeters in patients with different skin pigmentations. The authors summarise the pertinent literature and provide an overview of the potential clinical implications. The editorial closes by providing insights into how pulse oximetry could be improved to mitigate against such potential shortcomings.

Introduction

Pulse oximetry is a non-invasive optical technique used to assess arterial blood oxygen saturation (SpO₂) and is arguably one of the greatest advancements in patient monitoring in recent years. It entered clinical use in the 1980s, and since then has become the standard for oxygen saturation monitoring during anaesthesia, in the recovery room, and in intensive care units across the world ¹. It is attractive because of its simplicity, low cost, and ability to quickly detect hypoxaemia. Recently, pulse oximetry has been incorporated into wearable devices such as smartwatches. The widespread use of pulse oximetry in clinical and now consumer settings underlines the need to ensure it is as accurate and reliable as possible.

The accuracy of pulse oximeters has received renewed attention recently, fuelled by the need to identify hypoxaemia quickly and accurately during the progression of Covid-19. This is particularly true with patient-led pulse oximetry measurements in the home, which inform decisions as to whether to go to hospital. Several factors can impact the accuracy of pulse oximeters, including low perfusion, ambient light, motion artifact, nail polish, probe positioning, and irregular heart rhythms. However, it is the impact of skin pigmentation that has received greatest attention, with large-scale studies observing different levels of accuracy in people of different races and ethnicities. Recent research has found that pulse oximeters miss hypoxaemia more often in Black subjects than White subjects ², and that missed hypoxaemia is associated with poorer outcomes in hospital patients ³. In response, the UK Government has commissioned an independent review of the health impact of potential bias in medical devices ⁴, and the National

Institute for Health Research has issued a grant call for research into the ‘diagnostic accuracy of pulse oximeters’ ⁵.

In this editorial we summarise the pertinent literature relating to the accuracy of pulse oximeters **in patients with** different skin pigmentations and racial backgrounds. We then summarise the proposed routes to ensuring pulse oximetry provides accurate measurements for all.

Principles of pulse oximetry and its limitations

It is important to understand how pulse oximetry works to appreciate its limitations. Pulse oximeters estimate SpO₂ through the analysis of optical signals. Two or more optical pulse wave signals (photoplethysmograms (PPGs)), are created by illuminating a tissue bed (such as the finger) at different wavelengths of light and measuring the amount of light absorbed. Each signal exhibits a pulse wave: the amount of light received decreases and then increases with each heartbeat, as the tissue’s blood volume increases in systole and then decreases in diastole. The PPGs are obtained simultaneously at different wavelengths of light (typically red (660 nm) and infrared (IR) (940 nm)), so that one is influenced more strongly by oxygenated blood, and the other is influenced more strongly by deoxygenated blood. SpO₂ can then be estimated by comparing the normalised amplitudes of the pulses in each signal: the amplitude of the pulsatile (‘alternating current’, AC) component is divided by the average amount of light received (the ‘direct current’, DC component). The ratio of their normalised amplitudes (R) is then calculated (equation 1), which relates to blood oxygenation. Finally, R is converted to SpO₂ using an empirical relationship (equation 2) which is learnt through calibration studies.

$$R = \frac{AC_{red}/DC_{red}}{AC_{IR}/DC_{IR}} \quad (1) \quad SpO_2 = 110 - 25 \cdot R \quad (2)$$

This provides insight into several of the limitations of pulse oximetry. Firstly, there is an implicit assumption that the relationship between the ratio of pulse amplitudes (R) and SpO₂ is similar for all subjects. Secondly, low perfusion can result in smaller variations in blood volume, and therefore smaller pulse amplitudes, reducing signal quality. Thirdly, movement and changes in ambient light levels can corrupt the signals, making the analysis unreliable. Pressing research questions include: (i) is a single empirical relationship suitable for all skin pigmentations; and (ii) is the amount of light received, and therefore the signal quality, influenced by skin pigmentation? With this background we now consider the literature on this topic and potential routes to improve pulse oximeters.

Literature on pulse oximetry and skin pigmentation

We identified 15 studies which compared the accuracy of pulse oximeters in groups of different race or skin pigmentation, using arterial blood oxygen saturation derived from blood gas measurements (SaO₂) as a reference ^{2 3 6 7 8 9 10 11 12 13 14 15 16 17 18}. The evidence has built since the late 1980s and has been enhanced by several important studies since 2020. The key findings are now discussed.

There is substantial evidence that at least some pulse oximeters are less accurate in Black patients (and those of darker skin pigmentation such as Asian patients) than White patients. The inaccuracies in Black patients mostly result in overestimating SaO₂, particularly at lower oxygen saturations (such as at SaO₂ < 90% ⁹). This is supported by emerging evidence from the UK ¹⁹. However, the magnitudes of these inaccuracies are relatively small, such as a median difference in SaO₂ of 0.6% between Black and White patients with an SpO₂ of 90% (the threshold for making a diagnosis of hypoxaemia) ³. There is also evidence that pulse oximeter signal quality can be lower in patients with darker skin pigmentation ^{7 10}. Only three studies did not find a difference in accuracy, all of which could have been underpowered to detect a difference (with 4, 14 and 34 subjects in the smallest group in these studies) ^{6 10 13}.

The inaccuracies observed in Black patients, as well as other potential differences in their clinical condition and healthcare, may explain differences in clinical outcomes. Several studies have observed greater levels of hidden hypoxaemia in Black patients, where the true oxygen saturation was low enough to warrant clinical intervention, but this was hidden by erroneously high SpO₂ measurements ^{2 3 15}. In this context, the key consideration is the performance of pulse oximeters for identifying hypoxaemia (*i.e.*, performance at an SaO₂ of approximately 90%). It has been reported that performance for identifying hypoxaemia is significantly worse in Black patients than White patients: **areas under the receiver-operating-characteristic curve of 0.84 (95% CI, 0.81-0.87) and 0.89 (0.87-0.91) respectively** ². Furthermore, Black patients experience hidden hypoxaemia up to three times as frequently as White patients ², and that disease state is associated with the occurrence of hidden hypoxaemia ³. Evidence is now emerging that not only is hidden hypoxaemia more frequent in Black patients, but that it is also associated with poorer outcomes in hospital patients ³. However, it is not yet clear to what extent other factors contribute to the observed differences in accuracy and outcomes, such as those relating to low perfusion and health disparities. For instance, White patients have been observed to be more likely to receive reference blood gas measurements than other patients ³.

There is also evidence that the performance of pulse oximeters varies between brands and between probe types ^{11 12 14}. For instance, the SaO₂ level below which significant inaccuracies occur varies between pulse oximeters ^{11 12}. Furthermore, data collated by the Open Oximetry Project (www.openoximetry.org) demonstrates the differing

performance of inexpensive pulse oximeters, many of which are used globally. For instance, some have an accuracy (root mean square error) of >4%, while others have an accuracy of <2% (e.g., JPD-500E by SantaMedical and Cms-50d by Contec Medical System, respectively).

There are several limitations to the current evidence. Firstly, most studies used race to stratify patients, rather than objective measurements of skin pigmentation. Secondly, recent large-scale retrospective studies cannot be used to quantify the accuracy of pulse oximetry as they used SpO₂ and SaO₂ measurements taken up to 5 or 10 minutes apart^{2 3}. Thirdly, research has mostly focused on adults, despite the importance of SpO₂ measurements in children and newborns (although¹³ and¹⁶ are notable exceptions). Fourthly, it is often not clear whether low quality SpO₂ measurements were excluded from analyses, which could be caused by low perfusion levels and associated with skin pigmentation. Finally, it is not clear how relevant older research is to current practice, as pulse oximeters have been refined since these studies. These limitations provide much motivation for further research.

Clinical implications

We now consider how the clinical use of pulse oximetry could be impacted by the evidence that the accuracy of pulse oximeters is influenced by skin pigmentation.

Pulse oximeters are widely used to identify hypoxaemia, such as when administering anaesthesia, when triaging patients in the Emergency Department, and during postoperative recovery. During the Covid-19 pandemic pulse oximeters have been used in the home to identify severe, unrecognised hypoxaemia, a common feature of acute Covid-19 and a predictor of worse outcomes²⁰. There are three potential ways to address this in the short-term. Firstly, pulse oximeter models could be identified which are less susceptible to inaccuracies due to skin pigmentation. This requires prospective studies assessing multiple pulse oximeters in real-world settings¹². Secondly, different threshold SpO₂ levels could be used to identify hypoxaemia according to the patient's skin pigmentation³. However, such thresholds would likely need to be specific to individual pulse oximeter models. Furthermore, selecting a threshold according to a patient's race could lead to inappropriate choice of thresholds in some patients whose skin pigmentation differs from the 'average' for their racial group²¹. Thirdly, in the context of Covid-19 it has been proposed that if pulse oximeters are inaccurate in "patients with darker skin", then they could be used to monitor changes in SpO₂²² to provide automatic administration of oxygen based on saturations thresholds²³, or detect patients who may benefit from oxygen therapy²⁴. However, research indicates that pulse oximetry does not necessarily track changes in true oxygen saturation

accurately ²⁵. Therefore, any steps taken in the short-term are unlikely to fully resolve the issue.

Pulse oximeters are also used to identify signs of disease and to inform diagnosis. Pulse oximetry is a fundamental component of polysomnography tests to diagnose sleep apnoea and assess its severity. In this case a pulse oximeter is used to identify desaturations indicative of partial reduction or complete cessation of breathing. Inaccuracies in pulse oximetry could therefore influence the accuracy of sleep apnoea diagnosis ²⁶. Pulse oximetry is also often used to screen newborn babies to identify low blood oxygen saturations associated with critical congenital heart defects ²⁷. The UK National Screening Committee has suggested that further research is required to support the use of pulse oximetry in this context ²⁸, and it now appears important to assess the impact of skin pigmentation on its utility.

Engineering Perspective

One potential explanation for the inaccuracies observed in patients with darker skin is that the additional melanin in the skin could impact SpO₂ estimates independently of blood oxygen saturation ¹⁸. As described earlier, the estimation of SpO₂ involves calculating the ratio of the normalised amplitudes of two pulse waves measured using red and infrared light. Each normalised amplitude is calculated by dividing the pulse wave amplitude by the amount of light transmitted through the tissue. The presence of melanin at higher concentrations, as found in patients with darker skin, and at different saturation values might have an effect in the amount of light transmitted at both wavelengths (mainly the red), which could result in an erroneous R value and hence an incorrect SpO₂ value. Therefore, it might not be reasonable to assume a single relationship relates R to SpO₂ for all skin pigmentations (equation 2). Perhaps pulse oximeters could use different relationships for different skin pigmentations, requiring either automatic recognition of skin pigmentation, or selection by the operator ¹¹.

Conclusion

From the presented 15 studies, 80% have concluded that pulse oximeters are less accurate (overestimate true blood oxygen saturation) in subjects with darker skin. Awareness of this phenomenon has been heightened in the Covid-19 pandemic, where pulse oximeters have been used to identify hypoxaemia, contributing to treatment decisions. There is now evidence that hospital patients in whom hypoxaemia is not recognised by pulse oximetry experience worse outcomes. The impact of inaccuracies in pulse oximetry may also extend to other clinical scenarios, such as the diagnosis of sleep apnoea, and screening of new-born babies for congenital heart defects. Therefore, it is of utmost importance that this issue is addressed.

It has been proposed that the use of existing pulse oximeters could be adapted to account for the inaccuracies associated with skin pigmentation. However, we believe that the best solution is to refine the design of pulse oximeters to eliminate any bias associated with skin pigmentation.

Contribution of authors: PAK, RAH and PHC wrote the first draft together; and RAH and PHC performed the literature review together. KHS contributed to the latter drafts of the editorial and strongly supported the discussion on the clinical implications of pulse oximetry.

Acknowledgements

PHC acknowledges funding from the British Heart Foundation (BHF) through grant (FS/20/20/34626).

Conflict of interest statement: The authors declare that they have no conflict of interest.

References

1. Kyriacou P, Allen J, editors. Photoplethysmography: Technology, Signal Analysis, and Applications [Internet]. 1st Editio. Elsevier; 2021.
2. Sjoding MW, Dickson RP, Iwashyna TJ, Gay SE, Valley TS. Racial Bias in Pulse Oximetry Measurement. *N Engl J Med* 2020; **383**: 2477–8
3. Wong A-KI, Charpignon M, Kim H, et al. Analysis of Discrepancies Between Pulse Oximetry and Arterial Oxygen Saturation Measurements by Race and Ethnicity and Association With Organ Dysfunction and Mortality. *JAMA Netw open* 2021; **4**: e2131674
4. UK Government Department of Health and Social Care. Review launched into the health impact of potential bias in medical devices [Internet]. GOV.UK. 2021 [accessed 2021 Dec 16]. Available from: <https://www.gov.uk/government/news/review-launched-into-the-health-impact-of-potential-bias-in-medical-devices>
5. National Institute for Health Research. 21/608 Diagnostic accuracy of pulse oximeters [Internet]. nihr.ac.uk. 2021 [accessed 2022 Jan 17]. Available from: <https://www.nihr.ac.uk/documents/21608-diagnostic-accuracy-of-pulse-oximeters/29470>
6. Gabrielczyk MR, Buist RJ. Pulse oximetry and postoperative hypothermia. *Anaesthesia* 1988; **43**: 402–4
7. Ries AL, Prewitt LM, Johnson JJ. Skin Color and Ear Oximetry. *Chest* 1989; **96**: 287–90
8. Jubran A, Tobin MJ. Reliability of pulse oximetry in titrating supplemental oxygen therapy in ventilator-dependent patients. *Chest* 1990; **97**: 1420–5
9. Lee KH, Hui KP, Tan WC, Lim TK. Factors influencing pulse oximetry as compared to functional arterial saturation in multi-ethnic Singapore. *Singapore Med. J.* 1993. p. 385–7
10. Adler JN, Hughes LA, Vivilecchia R, Camargo CA. Effect of skin pigmentation on pulse oximetry accuracy in the Emergency Department. *Acad Emerg Med* 1998; **5**: 965–70
11. Bickler PE, Feiner JR, Severinghaus JW. Effects of skin pigmentation on pulse oximeter accuracy at low saturation. *Anesthesiology* 2005; **102**: 715–9
12. Feiner JR, Severinghaus JW, Bickler PE. Dark Skin Decreases the Accuracy of Pulse Oximeters at Low Oxygen Saturation: The Effects of Oximeter Probe Type and Gender. *Anesth Analg* 2007; **105**: S18–23
13. Foglia EE, Whyte RK, Chaudhary A, et al. The Effect of Skin Pigmentation on the Accuracy of Pulse Oximetry in Infants with Hypoxemia. *J Pediatr* 2017; **182**: 375-377.e2
14. Ebmeier SJ, Barker M, Bacon M, et al. A Two Centre Observational Study of Simultaneous Pulse Oximetry and Arterial Oxygen Saturation Recordings in Intensive Care Unit Patients. *Anaesth Intensive Care* 2018; **46**: 297–303

15. Valbuena VSM, Barbaro RP, Claar D, et al. Racial Bias in Pulse Oximetry Measurement Among Patients About to Undergo Extracorporeal Membrane Oxygenation in 2019-2020. *Chest* 2021; **[In Press]**
16. Vesoulis Z, Tims A, Lodhi H, Lalos N, Whitehead H. Racial discrepancy in pulse oximeter accuracy in preterm infants. *J Perinatol* Springer US 2021; **[In Press]**
17. Wiles MD, El-Nayal A, Elton G, et al. The effect of patient ethnicity on the accuracy of peripheral pulse oximetry in patients with COVID-19 pneumonitis: a single-centre, retrospective analysis. *Anaesthesia* 2021; **[In Press]**
18. Okunlola O, Lipnick MS, Batchelder P, Bernstein M, Feiner J, Bickler P. Pulse Oximeter Performance, Racial Inequity, and the Work Ahead. *Respir Care* 2021; respcare.09795
19. Bangash MN, Hodson J, Evison F, et al. Determining the impact of ethnicity on the accuracy of measurements of oxygen saturations. A retrospective cohort study [Preprint]. *medRxiv* 2021;
20. Greenhalgh T, Knight M, Inda-Kim M, Fulop NJ, Leach J, Vindrola-Padros C. Remote management of covid-19 using home pulse oximetry and virtual ward support. *BMJ* 2021; **372**: n677
21. Norton HL. Variation in pulse oximetry readings: melanin, not ethnicity, is the appropriate variable to use when investigating bias. *Anaesthesia* 2021; **[In Press]**
22. NHS. FAQs – COVID Oximetry @home and COVID virtual wards [Internet]. nhs.uk. 2021 [accessed 2021 Dec 16]. Available from: <https://www.england.nhs.uk/nhs-at-home/faqs-for-for-covid-virtual-wards-and-covid-oximetry-home/>
23. Michard F, Shelley K, L'Her E. COVID-19: Pulse oximeters in the spotlight. *J Clin Monit Comput* 2021; **35**: 11-14.
24. Roberts CM, Bugler JR, Melchor R, Hetzel MR, Spiro SG. Value of pulse oximetry in screening for long-term oxygen therapy requirement. *Eur Respir J*. 1993; **6**: 559-62.
25. Perkins GD, McAuley DF, Giles S, Routledge H, Gao F. Do changes in pulse oximeter oxygen saturation predict equivalent changes in arterial oxygen saturation? *Crit Care* 2003; **7**: R67
26. Hunasikatti M. Racial bias in accuracy of pulse oximetry and its impact on assessments of hypopnea and T90 in clinical studies. *J Clin Sleep Med* 2021; **17**: 1145
27. Brown S, Liyanage S, Mikrou P, Singh A, Ewer AK. Newborn pulse oximetry screening in the UK: a 2020 survey. *Lancet* Elsevier Ltd; 2020; **396**: 881
28. UK National Screening Committee. Newborn screening for critical congenital heart disease and significant noncardiac conditions related to hypoxaemia using pulse oximetry screening [Internet]. 2019. Available from: <https://www.gov.uk/government/consultations/pulse-oximetry-in-newborn-physical-screening>

